



					Supplier Technical Specification	Model	Brand	Country of Origin	
Item ID	No	Furniture / Detailed technical requirement			Catalogue page / evidence				
LOT 01 Operating theatres									
Anesthetist stool									
EQ-001	1								
	1	Height-adjustable clinical stool with stable five-leg base.							
	2	Seat shall have a seamless or sealed washable cover resistant to disinfectants.							
	3	Height adjustment shall be smooth with secure locking.							
	4	Proposed seat-height range approximately 450–650 mm.							
	5	Safe working load at least 120 kg.							
	6	Frame/base shall resist corrosion.							
	7	Provide castors or glides suitable for the receiving room floor.							
	8	Provide foot-operated adjustment where needed to preserve sterile-hand workflow.							
	9	Demonstrate stability and adjustment under the intended working load.							
	10	Supply a backrest and foot ring for prolonged seated anesthesia work.							
	11	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.							
	12	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.							
	13	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.							
	14	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.							
	15	All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.							



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origon
16		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
17		Inspect assembly, stability and surface condition at delivery before use.					
18		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
19		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
20		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
21		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
22		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

Closed sterile-instrument cabinets

EQ-002	2						
	1	Closed storage cabinet with smooth nonporous cleanable surfaces.					
	2	Corrosion-resistant construction; use stainless steel minimum grade 304 for sterile-instrument storage.					
	3	Provide at least three adjustable shelves.					
	4	Doors shall close securely and include a lock with two keys where access control is needed.					
	5	Provide stable feet or plinth that allows cleaning and prevents tipping.					
	6	Shelf surfaces shall not damage stored packs.					
	7	Proposed overall dimensions approximately 900 mm wide × 450 mm deep × 1,800 mm high.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
8		Provide a declared shelf-load capacity of at least 30 kg per shelf.					
9		Confirm cabinet dimensions and door swing against the room layout before fabrication.					
10		Medicines or instruments requiring controlled temperature shall not be stored on the assumption that this cabinet provides it.					
11		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
12		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
13		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
14		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
15		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
16		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
17		Inspect assembly, stability and surface condition at delivery before use.					
18		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
19		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
20		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
21		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
22		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

Electric obstetric operating table

EQ-005

5



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
1		Powered operating table suitable for cesarean and gynecological surgery.					
2		Electric or electrohydraulic adjustment with emergency backup repositioning.					
3		Safe working load at least 200 kg in declared operative positions.					
4		Adjustable height shall include approximately 700–1,000 mm.					
5		Provide Trendelenburg and reverse Trendelenburg of at least 25 degrees.					
6		Provide lateral tilt of at least 15 degrees to each side.					
7		Back section shall raise at least 60 degrees.					
8		Provide removable split leg sections and adjustable head section.					
9		Radiolucent tabletop with disinfectant-resistant, antistatic pressure-distributing mattress.					
10		Frame and exposed surfaces shall resist corrosion and cleaning chemicals.					
11		Provide secure floor locking or base braking.					
12		Supply two arm boards, anesthesia screen, patient straps and one pair of padded lithotomy supports with clamps.					
13		Provide compatible accessory side rails and controls protected from accidental activation.					
14		Demonstrate safe movement under the declared load and emergency operation.					
15		This item is allocated to Khanfar only; confirm a functional operating table is available at Al-Nashmah.					
16		The table shall support at least 200 kg safe working load in every manufacturer-approved clinical position, including the supplied accessories.					
17		The specified assembly shall accommodate the intended patient/load and fit the approved installation or storage space; submit dimensioned manufacturer evidence before award.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
18		The specified assembly shall accommodate the intended patient/load and fit the approved installation or storage space; submit dimensioned manufacturer evidence before award.					
19		The supplied configuration shall include longitudinal slide availability and limits if offered as an operational feature, with manufacturer evidence and acceptance demonstration.					
20		The specified assembly shall accommodate the intended patient/load and fit the approved installation or storage space; submit dimensioned manufacturer evidence before award.					
21		Provide accessible emergency stop or equivalent safe motion interruption.					
22		Provide a protected hand controller and secure controller storage.					
23		Emergency battery shall permit at least 10 full table-adjustment cycles under a representative operating load after mains failure, including a low-battery warning.					
24		A manual or battery-backed emergency method shall permit safe lowering to patient transfer height during mains or control failure.					
25		The floor-lock indication shall be visible or verifiable before surgery.					
26		Accessory clamps shall match the table rail dimensions and lock securely.					
27		Lithotomy supports shall have declared load limits and padded patient supports.					
28		Mattress sections shall fit securely and allow approved disinfection.					
29		The supplied table shall permit clinically usable intraoperative imaging access under the patient top without interference from the base at the intended positions.					
30		Supply a complete accessory list with individual quantities and catalogue references.					
31		All relevant components shall withstand the manufacturer-validated cleaning, disinfection or sterilization method required for their specified clinical reuse; include validated instructions.					
32		Demonstrate movement limits and stability with a safe representative test load.					
33		Train users in positioning, pinch-point avoidance and emergency release.					
34		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	35	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	36	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	37	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Height-adjustable surgeon stool							
EQ-007	7						
	1	Height-adjustable clinical stool with stable five-leg base.					
	2	Seat shall have a seamless or sealed washable cover resistant to disinfectants.					
	3	Height adjustment shall be smooth with secure locking.					
	4	Proposed seat-height range approximately 450–650 mm.					
	5	Safe working load at least 120 kg.					
	6	Frame/base shall resist corrosion.					
	7	Provide castors or glides suitable for the receiving room floor.					
	8	Provide foot-operated adjustment where needed to preserve sterile-hand workflow.					
	9	Demonstrate stability and adjustment under the intended working load.					
	10	Provide a compact seat and unobstructed access suitable for the operating surgeon; confirm backrest preference with the users.					
	11	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	12	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
13		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
14		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
15		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
16		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
17		Inspect assembly, stability and surface condition at delivery before use.					
18		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
19		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
20		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
21		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
22		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					

TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply

Instrument preparation table							
EQ-008	8						
1		Stainless-steel instrument preparation table for a clean processing area.					
2		Minimum grade 304 stainless-steel top with reinforced underside.					
3		Proposed worktop dimensions approximately 1,500 mm × 700 mm and height 850–900 mm.					
4		Worktop thickness at least 1.2 mm.					
5		Provide smooth welded joints and rounded exposed edges.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origon
6		Provide adjustable feet or braked castors as approved by the hospital.					
7		Provide lower shelf where it does not interfere with seated/standing work requirements.					
8		Declared evenly distributed load at least 100 kg.					
9		Surfaces shall tolerate routine disinfection and remain free of fluid traps.					
10		Confirm dimensions and clean-area workflow before fabrication.					
11		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
12		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
13		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
14		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
15		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
16		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
17		Inspect assembly, stability and surface condition at delivery before use.					
18		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
19		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
20		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
21		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
22		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
EQ-010	10	Lockable anesthesia trolley					
	1	Hospital mobile trolley with corrosion-resistant frame and smooth disinfectant-resistant surfaces.					
	2	Provide four castors with at least two brakes.					
	3	Provide rounded edges, push handle and stable construction.					
	4	Supply tray/shelf retaining edges appropriate to the contents.					
	5	The item shall fit the receiving room and route of access; its load rating shall meet the item-specific minimum stated above, with dimensioned manufacturer drawing submitted for verification.					
	6	Proposed footprint approximately 600–800 mm wide × 400–550 mm deep; confirm room and doorway fit.					
	7	Components shall tolerate repeated cleaning without trapping liquid or debris.					
	8	Demonstrate drawer/door function where fitted and loaded trolley stability.					
	9	Anesthesia configuration with at least five drawers and adjustable dividers.					
	10	Provide central lock and two keys.					
	11	Provide recessed work surface and brackets for consumable containers.					
	12	Drugs and anesthesia machine are excluded from this trolley line.					
	13	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	14	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	15	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
	16	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origon
17		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
18		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
19		Inspect assembly, stability and surface condition at delivery before use.					
20		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
21		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
22		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
23		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
24		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

Mobile IV pole							
EQ-011	11						
	1	Mobile IV pole with corrosion-resistant stainless-steel upright.					
	2	Telescopic height adjustment approximately 1,200–2,000 mm.					
	3	Provide at least four IV-bag hooks.					
	4	Stable five-leg wheeled base.					
	5	Provide positive height-locking mechanism.					
	6	Rated load at least 2 kg per hook, subject to the declared overall pole rating.					
	7	Provide at least two braked castors.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origon
	8	Pole shall accept compatible infusion-pump clamps without instability.					
	9	Surfaces shall be smooth and easy to disinfect.					
	10	Demonstrate stability with the intended pump/bag configuration.					
	11	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	12	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	13	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
	14	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
	15	All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
	16	Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
	17	Inspect assembly, stability and surface condition at delivery before use.					
	18	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	19	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	20	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	21	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	22	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

Mobile Mayo table



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
1		Mobile Mayo instrument table with removable stainless-steel tray.					
2		Tray dimensions approximately 500–600 mm × 350–450 mm.					
3		Height adjustment approximately 800–1,300 mm.					
4		Provide a secure height lock without unintended lowering.					
5		Base shall slide beneath/alongside the operating table without collision.					
6		Provide stable castors with braking or equivalent secure positioning.					
7		All exposed metal shall be corrosion resistant, minimum grade 304 for the tray.					
8		Tray edges shall be smooth and raised to retain instruments.					
9		Safe tray load at least 10 kg evenly distributed.					
10		Demonstrate stability at maximum height with the declared load.					
11		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
12		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
13		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
14		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
15		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
16		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
17		Inspect assembly, stability and surface condition at delivery before use.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	18	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	19	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	20	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	21	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	22	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Patient step stool							
EQ-013	13						
	1	Patient step stool with non-slip top surface.					
	2	Safe working load at least 150 kg.					
	3	Corrosion-resistant frame with smooth rounded edges.					
	4	Provide rubber non-slip feet.					
	5	Proposed top size approximately 450 mm × 300 mm and height 200–250 mm.					
	6	Step shall remain stable under off-centre loading.					
	7	Top surface shall be washable and securely fixed.					
	8	Provide no sharp edges or exposed fixing ends.					
	9	Confirm height against the examination couch or operating table approach.					
	10	Demonstrate stability and supply declared load-test information.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
	11	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	12	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	13	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
	14	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
	15	All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
	16	Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
	17	Inspect assembly, stability and surface condition at delivery before use.					
	18	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	19	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	20	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	21	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	22	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Patient transfer board							
EQ-014	14						
	1	Patient lateral-transfer board suitable for bed-to-stretcher transfer.					
	2	Smooth low-friction surface with rounded edges.					
	3	Safe working load at least 150 kg under the manufacturer's specified supported-transfer conditions.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
4		Board length approximately 1,700–1,900 mm or a validated folding equivalent.					
5		Provide handholds or a safe handling arrangement.					
6		Material shall resist cleaning agents and bodily fluids.					
7		Supply a compatible low-friction cover/sleeve where needed for the design.					
8		The supplied configuration shall include allowable transfer gap and required number of trained operators as an operational feature, with manufacturer evidence and acceptance demonstration.					
9		Board shall not be represented as an unsupported stretcher.					
10		Demonstrate safe transfer technique using a training load.					
11		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
12		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
13		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
14		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
15		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
16		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
17		Inspect assembly, stability and surface condition at delivery before use.					
18		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
19		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
20		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	21	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	22	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Patient transfer stretcher							
EQ-015	15						
	1	Hospital patient-transfer stretcher with adjustable backrest and height.					
	2	Safe working load at least 180 kg.					
	3	Patient platform length at least 1,850 mm and width at least 550 mm.					
	4	Provide secure folding side rails and mattress-retaining features.					
	5	Supply waterproof cleanable transfer mattress.					
	6	Provide four large castors with effective brakes and directional steering.					
	7	Provide push handles and bumpers.					
	8	Supply IV pole and oxygen-cylinder holder.					
	9	Frame shall resist corrosion and routine hospital disinfectants.					
	10	Demonstrate transfer height, stability, rail locking and braking under declared load.					
	11	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	12	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	13	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
14		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
15		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
16		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
17		Inspect assembly, stability and surface condition at delivery before use.					
18		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
19		Patient support and mattress shall fit together without unsafe gaps and accommodate the patient group and load capacity stated for this item; provide a dimensioned drawing.					
20		Movable sections shall have protected mechanisms and safe clearances to prevent patient or operator entrapment during normal adjustment.					
21		All mobile equipment shall have hospital-grade casters with effective brakes on at least two wheels and shall roll safely over typical indoor door thresholds.					
22		Upholstery or mattress covers shall be fluid resistant, intact and readily disinfected.					
23		Patient transfer shall be possible using the supplied handles/rails and standard clinical staff technique without removing safety components.					
24		Demonstrate safe entry/exit or transfer and locking under the declared working conditions.					
25		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
26		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
27		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
28		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

Surgical instrument trolley

EQ-017 17



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
1		Hospital mobile trolley with corrosion-resistant frame and smooth disinfectant-resistant surfaces.					
2		Provide four castors with at least two brakes.					
3		Provide rounded edges, push handle and stable construction.					
4		Supply tray/shelf retaining edges appropriate to the contents.					
5		The item shall fit the receiving room and route of access; its load rating shall meet the item-specific minimum stated above, with dimensioned manufacturer drawing submitted for verification.					
6		Proposed footprint approximately 600–800 mm wide × 400–550 mm deep; confirm room and doorway fit.					
7		Components shall tolerate repeated cleaning without trapping liquid or debris.					
8		Demonstrate drawer/door function where fitted and loaded trolley stability.					
9		Surgical instrument configuration with two removable stainless-steel trays/shelves.					
10		Provide sufficient clearance between shelves for covered instrument trays.					
11		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
12		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
13		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
14		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
15		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
16		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
17		Inspect assembly, stability and surface condition at delivery before use.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
18		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
19		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
20		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
21		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
22		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

LOT 03 Recovery and preparation**Anesthesia trolley, lockable**

EQ-027	27						
1		Hospital mobile trolley with corrosion-resistant frame and smooth disinfectant-resistant surfaces.					
2		Provide four castors with at least two brakes.					
3		Provide rounded edges, push handle and stable construction.					
4		Supply tray/shelf retaining edges appropriate to the contents.					
5		The item shall fit the receiving room and route of access; its load rating shall meet the item-specific minimum stated above, with dimensioned manufacturer drawing submitted for verification.					
6		Proposed footprint approximately 600–800 mm wide × 400–550 mm deep; confirm room and doorway fit.					
7		Components shall tolerate repeated cleaning without trapping liquid or debris.					
8		Demonstrate drawer/door function where fitted and loaded trolley stability.					
9		Anesthesia configuration with at least five drawers and adjustable dividers.					
10		Provide central lock and two keys.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
	11	Provide recessed work surface and brackets for consumable containers.					
	12	Drugs and anesthesia machine are excluded from this trolley line.					
	13	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	14	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	15	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
	16	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
	17	All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
	18	Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
	19	Inspect assembly, stability and surface condition at delivery before use.					
	20	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	21	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	22	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	23	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	24	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Closed clean-instrument trolley							
EQ-028	28						
	1	Hospital mobile trolley with corrosion-resistant frame and smooth disinfectant-resistant surfaces.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	2	Provide four castors with at least two brakes.					
	3	Provide rounded edges, push handle and stable construction.					
	4	Supply tray/shelf retaining edges appropriate to the contents.					
	5	The item shall fit the receiving room and route of access; its load rating shall meet the item-specific minimum stated above, with dimensioned manufacturer drawing submitted for verification.					
	6	Proposed footprint approximately 600–800 mm wide × 400–550 mm deep; confirm room and doorway fit.					
	7	Components shall tolerate repeated cleaning without trapping liquid or debris.					
	8	Demonstrate drawer/door function where fitted and loaded trolley stability.					
	9	Closed clean-instrument transport configuration with smooth hinged doors.					
	10	Provide removable internal shelves and secure closure during movement.					
	11	Enclosure shall protect cleaned/packed instruments from dust and splashes.					
	12	Provide a clean-use identification label; do not use the same trolley for dirty instruments.					
	13	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	14	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	15	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
	16	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
	17	All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
	18	Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
	19	Inspect assembly, stability and surface condition at delivery before use.					
	20	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	21	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	22	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	23	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	24	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Recovery bed with side rails							
EQ-033	33						
	1	Adult hospital bed with corrosion-resistant steel frame and smooth disinfectant-resistant finish.					
	2	Safe working load at least 180 kg, including patient and accessories.					
	3	Provide adjustable backrest and knee/leg section.					
	4	Provide secure, collapsible side rails with safe gaps and protected pinch points.					
	5	Removable head and foot boards for clinical access.					
	6	Supply a pressure-distributing mattress with removable waterproof cover and sealed or protected seams.					
	7	Bed and mattress dimensions shall be compatible; declare overall dimensions and usable sleeping area.					
	8	Four castors with at least two effective brakes.					
	9	Supply one IV pole and urine-bag hooks.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
10		Demonstrate positioning, rail locking, brake function and safe working load documentation.					
11		Recovery bed shall provide height adjustment and a rapid flat position.					
12		Backrest shall reach at least 60 degrees.					
13		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
14		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
15		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
16		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
17		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
18		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
19		Inspect assembly, stability and surface condition at delivery before use.					
20		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
21		Patient support and mattress shall fit together without unsafe gaps and accommodate the patient group and load capacity stated for this item; provide a dimensioned drawing.					
22		Movable sections shall have protected mechanisms and safe clearances to prevent patient or operator entrapment during normal adjustment.					
23		All mobile equipment shall have hospital-grade casters with effective brakes on at least two wheels and shall roll safely over typical indoor door thresholds.					
24		Upholstery or mattress covers shall be fluid resistant, intact and readily disinfected.					
25		Patient transfer shall be possible using the supplied handles/rails and standard clinical staff technique without removing safety components.					
26		Demonstrate safe entry/exit or transfer and locking under the declared working conditions.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	27	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	28	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	29	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	30	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

LOT 04 Intensive care and high dependency

Five-function electric ICU bed							
EQ-037	37						
	1	Electric ICU bed with five powered functions: height, backrest, knee section, Trendelenburg and reverse Trendelenburg.					
	2	Safe working load at least 200 kg including accessories.					
	3	Backrest adjustment to at least 70 degrees.					
	4	Trendelenburg and reverse Trendelenburg adjustment at least 12 degrees in each direction.					
	5	Bed-height adjustment range shall include approximately 450–750 mm, excluding mattress.					
	6	Provide manual emergency CPR release and a flat resuscitation position.					
	7	Battery backup shall allow safe repositioning during a mains failure; supplier shall state tested cycle capacity.					
	8	Provide lockable patient controls and nurse override.					
	9	Split side rails shall lock securely and meet applicable entrapment and mechanical safety requirements.					
	10	Four heavy-duty castors with central braking and directional steering.					
	11	Supply pressure-redistributing ICU mattress with waterproof cleanable cover.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
12		Provide IV pole, urine-bag hooks and removable head/foot boards.					
13		Demonstrate all five functions, battery operation, CPR release and braking before acceptance.					
14		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
15		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
16		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
17		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
18		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
19		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
20		Inspect assembly, stability and surface condition at delivery before use.					
21		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
22		Patient support and mattress shall fit together without unsafe gaps and accommodate the patient group and load capacity stated for this item; provide a dimensioned drawing.					
23		Movable sections shall have protected mechanisms and safe clearances to prevent patient or operator entrapment during normal adjustment.					
24		All mobile equipment shall have hospital-grade casters with effective brakes on at least two wheels and shall roll safely over typical indoor door thresholds.					
25		Upholstery or mattress covers shall be fluid resistant, intact and readily disinfected.					
26		Patient transfer shall be possible using the supplied handles/rails and standard clinical staff technique without removing safety components.					
27		Demonstrate safe entry/exit or transfer and locking under the declared working conditions.					
28		All functions specified for the bed type (backrest, knee, height and tilt as applicable) shall be independently operable through their full manufacturer-rated range, without binding under rated load.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
	29	The specified assembly shall accommodate the intended patient/load and fit the approved installation or storage space; submit dimensioned manufacturer evidence before award.					
	30	Provide protected controls or crank storage to limit accidental activation.					
	31	The bed shall have a quick manual CPR release to flatten the backrest without mains power.					
	32	Provide secure side-rail latches where rails are included.					
	33	The described component shall be user-serviceable or replaceable by the authorized service team, with the necessary service information and five-year availability in Yemen.					
	34	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	35	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	36	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	37	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
ICU crash cart							
EQ-038	38						
	1	Mobile emergency trolley with at least five drawers and a washable work surface.					
	2	Central drawer locking or tamper-evident breakaway sealing system.					
	3	Clearly labelled drawer compartments for organized emergency supplies.					
	4	Raised worktop edges and smooth disinfectant-resistant surfaces.					
	5	Four castors with at least two brakes.					
	6	Provide a defibrillator shelf compatible with the separately scheduled defibrillator or the existing unit.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	7	Provide a rigid CPR board.					
	8	Provide IV pole, oxygen-cylinder holder and sharps-container bracket.					
	9	Overall dimensions and loaded stability shall suit the receiving ward and doorways.					
	10	Supply 20 numbered tamper seals if the offered locking system uses them.					
	11	Trolley supply excludes medicines and separately scheduled medical devices; quote all trolley accessories in the unit price.					
	12	Demonstrate drawer access, brakes and accessory retention under the declared safe load.					
	13	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	14	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	15	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
	16	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
	17	All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
	18	Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
	19	Inspect assembly, stability and surface condition at delivery before use.					
	20	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	21	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	22	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	23	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	24	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
		TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply					
LOT 05 Maternity and obstetrics							
IV pole, privacy curtain and bedside step							
EQ-050	50						
	1	Each supply set shall include one mobile IV pole, one complete privacy-curtain assembly and one bedside step.					
	2	IV pole shall have a stable five-leg base, at least two hooks and height adjustment approximately 1,200–2,000 mm.					
	3	Privacy curtain shall be opaque, washable and compatible with hospital infection-control practices.					
	4	Include curtain track, gliders, end stops and mounting hardware for one bed space.					
	5	Track length and shape shall be measured and approved before manufacture.					
	6	Bedside step shall have a non-slip top and feet with safe working load at least 150 kg.					
	7	All metal components shall resist corrosion and cleaning agents.					
	8	Provide a complete installation drawing for the curtain assembly.					
	9	Confirm whether existing tracks can be used; do not omit required fixing materials from the quoted set.					
	10	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	11	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	12	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
	13	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
	14	All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	15	Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
	16	Inspect assembly, stability and surface condition at delivery before use.					
	17	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	18	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	19	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	20	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	21	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Multi-position obstetric delivery bed							
EQ-053	53						
	1	Multi-position obstetric bed for labour, delivery and immediate recovery.					
	2	Three-section mattress platform with removable/retractable leg section and perineal access.					
	3	Adjustable bed height and backrest angle; backrest shall reach at least 60 degrees.					
	4	Provide Trendelenburg and reverse Trendelenburg positions of at least 15 degrees.					
	5	Safe working load at least 180 kg.					
	6	Electrically powered adjustment with emergency manual or battery-supported repositioning.					
	7	Supply mattress with washable waterproof cover and secure attachment.					
	8	Provide retractable side rails and maternal hand grips.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	9	Provide two adjustable padded leg supports/stirrups.					
	10	Supply removable stainless-steel fluid collection bowl/tray.					
	11	Supply IV pole, urine-bag hooks and compatible bedside step.					
	12	Four heavy-duty castors with effective braking.					
	13	The specified assembly shall accommodate the intended patient/load and fit the approved installation or storage space; submit dimensioned manufacturer evidence before award.					
	14	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	15	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	16	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
	17	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
	18	All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
	19	Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
	20	Inspect assembly, stability and surface condition at delivery before use.					
	21	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	22	Patient support and mattress shall fit together without unsafe gaps and accommodate the patient group and load capacity stated for this item; provide a dimensioned drawing.					
	23	Movable sections shall have protected mechanisms and safe clearances to prevent patient or operator entrapment during normal adjustment.					
	24	All mobile equipment shall have hospital-grade casters with effective brakes on at least two wheels and shall roll safely over typical indoor door thresholds.					
	25	Upholstery or mattress covers shall be fluid resistant, intact and readily disinfected.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
26		Patient transfer shall be possible using the supplied handles/rails and standard clinical staff technique without removing safety components.					
27		Demonstrate safe entry/exit or transfer and locking under the declared working conditions.					
28		All functions specified for the bed type (backrest, knee, height and tilt as applicable) shall be independently operable through their full manufacturer-rated range, without binding under rated load.					
29		The specified assembly shall accommodate the intended patient/load and fit the approved installation or storage space; submit dimensioned manufacturer evidence before award.					
30		Provide protected controls or crank storage to limit accidental activation.					
31		The bed shall have a quick manual CPR release to flatten the backrest without mains power.					
32		Provide secure side-rail latches where rails are included.					
33		The described component shall be user-serviceable or replaceable by the authorized service team, with the necessary service information and five-year availability in Yemen.					
34		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
35		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
36		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
37		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

LOT 06 Neonatal care

Gynecological examination couch							
EQ-059	59						
1		Clinical examination couch with rigid corrosion-resistant frame.					
2		Patient surface shall have a seamless or sealed waterproof cover resistant to routine disinfectants.					
3		Provide padded mattress or upholstery securely attached to the frame.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
4		All edges and corners shall be smooth.					
5		Provide stable non-slip feet or lockable castors if a mobile design is offered.					
6		The item shall fit the receiving room and route of access; its load rating shall meet the item-specific minimum stated above, with dimensioned manufacturer drawing submitted for verification.					
7		Supply paper-roll holder where compatible with the intended use.					
8		Demonstrate stability and cleaning access.					
9		Adult patient surface with adjustable backrest and gynecological perineal cutout.					
10		Safe working load at least 180 kg.					
11		Supply two adjustable padded leg supports and a removable stainless-steel collection bowl.					
12		Supply a non-slip patient step.					
13		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
14		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
15		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
16		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
17		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
18		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
19		Inspect assembly, stability and surface condition at delivery before use.					
20		Demonstrate all adjustments, locks and brakes fitted to the offered item.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	21	Patient support and mattress shall fit together without unsafe gaps and accommodate the patient group and load capacity stated for this item; provide a dimensioned drawing.					
	22	Movable sections shall have protected mechanisms and safe clearances to prevent patient or operator entrapment during normal adjustment.					
	23	All mobile equipment shall have hospital-grade casters with effective brakes on at least two wheels and shall roll safely over typical indoor door thresholds.					
	24	Upholstery or mattress covers shall be fluid resistant, intact and readily disinfected.					
	25	Patient transfer shall be possible using the supplied handles/rails and standard clinical staff technique without removing safety components.					
	26	Demonstrate safe entry/exit or transfer and locking under the declared working conditions.					
	27	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	28	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	29	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	30	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Infant cot/bassinet							
EQ-060	60						
	1	Mobile infant cot/bassinet with transparent removable infant compartment.					
	2	Smooth rounded surfaces allowing observation and cleaning.					
	3	Supply a close-fitting firm infant mattress with waterproof cleanable cover.					
	4	Provide a stable corrosion-resistant stand.					
	5	Provide four castors with brakes.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
6		Provide controlled tilt if included, with positive position locking.					
7		Declare safe infant load; proposed minimum 10 kg.					
8		Provide no gaps or projections that could trap or injure the infant.					
9		Supply lower storage shelf if it does not compromise stability.					
10		Demonstrate brakes, stability and cleaning access.					
11		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
12		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
13		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
14		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
15		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
16		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
17		Inspect assembly, stability and surface condition at delivery before use.					
18		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
19		Patient support and mattress shall fit together without unsafe gaps and accommodate the patient group and load capacity stated for this item; provide a dimensioned drawing.					
20		Movable sections shall have protected mechanisms and safe clearances to prevent patient or operator entrapment during normal adjustment.					
21		All mobile equipment shall have hospital-grade casters with effective brakes on at least two wheels and shall roll safely over typical indoor door thresholds.					
22		Upholstery or mattress covers shall be fluid resistant, intact and readily disinfected.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	23	Patient transfer shall be possible using the supplied handles/rails and standard clinical staff technique without removing safety components.					
	24	Demonstrate safe entry/exit or transfer and locking under the declared working conditions.					
	25	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	26	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	27	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	28	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Neonatal resuscitation trolley							
EQ-070	70						
	1	Hospital mobile trolley with corrosion-resistant frame and smooth disinfectant-resistant surfaces.					
	2	Provide four castors with at least two brakes.					
	3	Provide rounded edges, push handle and stable construction.					
	4	Supply tray/shelf retaining edges appropriate to the contents.					
	5	The item shall fit the receiving room and route of access; its load rating shall meet the item-specific minimum stated above, with dimensioned manufacturer drawing submitted for verification.					
	6	Proposed footprint approximately 600–800 mm wide × 400–550 mm deep; confirm room and doorway fit.					
	7	Components shall tolerate repeated cleaning without trapping liquid or debris.					
	8	Demonstrate drawer/door function where fitted and loaded trolley stability.					
	9	Neonatal resuscitation storage configuration with at least four drawers and labelled compartments.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
10		Provide holders for neonatal airway supplies, oxygen tubing and suction accessories.					
11		This trolley stores resuscitation supplies; a heated neonatal patient platform is covered by the separately scheduled neonatal radiant warmer.					
12		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
13		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
14		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
15		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
16		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
17		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
18		Inspect assembly, stability and surface condition at delivery before use.					
19		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
20		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
21		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
22		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
23		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

LOT 07 Emergency outpatient and diagnostics

Emergency crash cart							
EQ-081	81						
1		Mobile emergency trolley with at least five drawers and a washable work surface.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	2	Central drawer locking or tamper-evident breakaway sealing system.					
	3	Clearly labelled drawer compartments for organized emergency supplies.					
	4	Raised worktop edges and smooth disinfectant-resistant surfaces.					
	5	Four castors with at least two brakes.					
	6	Provide a defibrillator shelf compatible with the separately scheduled defibrillator or the existing unit.					
	7	Provide a rigid CPR board.					
	8	Provide IV pole, oxygen-cylinder holder and sharps-container bracket.					
	9	Overall dimensions and loaded stability shall suit the receiving ward and doorways.					
	10	Supply 20 numbered tamper seals if the offered locking system uses them.					
	11	Trolley supply excludes medicines and separately scheduled medical devices; quote all trolley accessories in the unit price.					
	12	Demonstrate drawer access, brakes and accessory retention under the declared safe load.					
	13	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	14	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	15	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
	16	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
	17	All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
	18	Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	19	Inspect assembly, stability and surface condition at delivery before use.					
	20	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	21	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	22	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	23	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	24	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Emergency resuscitation bed							
EQ-082	82						
	1	Adult hospital bed with corrosion-resistant steel frame and smooth disinfectant-resistant finish.					
	2	Safe working load at least 180 kg, including patient and accessories.					
	3	Provide adjustable backrest and knee/leg section.					
	4	Provide secure, collapsible side rails with safe gaps and protected pinch points.					
	5	Removable head and foot boards for clinical access.					
	6	Supply a pressure-distributing mattress with removable waterproof cover and sealed or protected seams.					
	7	Bed and mattress dimensions shall be compatible; declare overall dimensions and usable sleeping area.					
	8	Four castors with at least two effective brakes.					
	9	Supply one IV pole and urine-bag hooks.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
10		Demonstrate positioning, rail locking, brake function and safe working load documentation.					
11		Resuscitation bed shall include a rapid flat/CPR position and height adjustment.					
12		Provide access around the bed for resuscitation and easy patient transfer.					
13		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
14		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
15		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
16		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
17		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
18		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
19		Inspect assembly, stability and surface condition at delivery before use.					
20		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
21		Patient support and mattress shall fit together without unsafe gaps and accommodate the patient group and load capacity stated for this item; provide a dimensioned drawing.					
22		Movable sections shall have protected mechanisms and safe clearances to prevent patient or operator entrapment during normal adjustment.					
23		All mobile equipment shall have hospital-grade casters with effective brakes on at least two wheels and shall roll safely over typical indoor door thresholds.					
24		Upholstery or mattress covers shall be fluid resistant, intact and readily disinfected.					
25		Patient transfer shall be possible using the supplied handles/rails and standard clinical staff technique without removing safety components.					
26		Demonstrate safe entry/exit or transfer and locking under the declared working conditions.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
	27	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	28	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	29	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	30	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
Infant Examination Table							
EQ-086	86						
	1	Clinical examination couch with rigid corrosion-resistant frame.					
	2	Patient surface shall have a seamless or sealed waterproof cover resistant to routine disinfectants.					
	3	Provide padded mattress or upholstery securely attached to the frame.					
	4	All edges and corners shall be smooth.					
	5	Provide stable non-slip feet or lockable castors if a mobile design is offered.					
	6	The item shall fit the receiving room and route of access; its load rating shall meet the item-specific minimum stated above, with dimensioned manufacturer drawing submitted for verification.					
	7	Supply paper-roll holder where compatible with the intended use.					
	8	Demonstrate stability and cleaning access.					
	9	Infant patient surface approximately 900–1,200 mm long and 600–700 mm wide.					
	10	Safe working load at least 30 kg.					
	11	Provide raised protective edges appropriate to infant examination and a secure padded changing surface.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
12		Provide under-surface storage clear of cleaning areas.					
13		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
14		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
15		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
16		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
17		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
18		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
19		Inspect assembly, stability and surface condition at delivery before use.					
20		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
21		Patient support and mattress shall fit together without unsafe gaps and accommodate the patient group and load capacity stated for this item; provide a dimensioned drawing.					
22		Movable sections shall have protected mechanisms and safe clearances to prevent patient or operator entrapment during normal adjustment.					
23		All mobile equipment shall have hospital-grade casters with effective brakes on at least two wheels and shall roll safely over typical indoor door thresholds.					
24		Upholstery or mattress covers shall be fluid resistant, intact and readily disinfected.					
25		Patient transfer shall be possible using the supplied handles/rails and standard clinical staff technique without removing safety components.					
26		Demonstrate safe entry/exit or transfer and locking under the declared working conditions.					
27		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
28		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	29	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	30	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Dressing trolley							
EQ-094	94						
	1	Hospital mobile trolley with corrosion-resistant frame and smooth disinfectant-resistant surfaces.					
	2	Provide four castors with at least two brakes.					
	3	Provide rounded edges, push handle and stable construction.					
	4	Supply tray/shelf retaining edges appropriate to the contents.					
	5	The item shall fit the receiving room and route of access; its load rating shall meet the item-specific minimum stated above, with dimensioned manufacturer drawing submitted for verification.					
	6	Proposed footprint approximately 600–800 mm wide × 400–550 mm deep; confirm room and doorway fit.					
	7	Components shall tolerate repeated cleaning without trapping liquid or debris.					
	8	Demonstrate drawer/door function where fitted and loaded trolley stability.					
	9	Dressing configuration with two stainless-steel shelves.					
	10	Include one bowl, one kidney dish and one waste-container holder.					
	11	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	12	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	13	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origon
	14	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
	15	All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
	16	Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
	17	Inspect assembly, stability and surface condition at delivery before use.					
	18	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	19	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	20	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	21	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	22	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Medicine trolley							
EQ-095	95						
	1	Hospital mobile trolley with corrosion-resistant frame and smooth disinfectant-resistant surfaces.					
	2	Provide four castors with at least two brakes.					
	3	Provide rounded edges, push handle and stable construction.					
	4	Supply tray/shelf retaining edges appropriate to the contents.					
	5	The item shall fit the receiving room and route of access; its load rating shall meet the item-specific minimum stated above, with dimensioned manufacturer drawing submitted for verification.					
	6	Proposed footprint approximately 600–800 mm wide × 400–550 mm deep; confirm room and doorway fit.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origon
7		Components shall tolerate repeated cleaning without trapping liquid or debris.					
8		Demonstrate drawer/door function where fitted and loaded trolley stability.					
9		Medication configuration with at least four lockable drawers and dividers.					
10		Provide central locking with two keys.					
11		Include a washable medicine-preparation top.					
12		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
13		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
14		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
15		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
16		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
17		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
18		Inspect assembly, stability and surface condition at delivery before use.					
19		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
20		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
21		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
22		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
23		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
EQ-096	96	Mobile overbed table					
	1	Height-adjustable mobile overbed table.					
	2	Tabletop approximately 750–900 mm long and 350–450 mm deep.					
	3	Height-adjustment range shall cover approximately 700–1,000 mm.					
	4	Top shall be nonporous, moisture resistant and easy to disinfect.					
	5	Provide raised retaining edge where appropriate.					
	6	Base shall fit under the intended hospital bed without interfering with bed movement.					
	7	Provide castors with at least two brakes.					
	8	Rated tabletop load at least 10 kg evenly distributed.					
	9	Height lock shall prevent unintended lowering.					
	10	Demonstrate compatibility with the beds at the receiving hospital.					
	11	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	12	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	13	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
	14	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
	15	All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
	16	Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
	17	Inspect assembly, stability and surface condition at delivery before use.					
	18	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	19	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	20	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	21	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	22	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Patient commode chair							
EQ-097	97						
	1	Adult bedside commode chair with corrosion-resistant frame.					
	2	Safe working load at least 120 kg.					
	3	Seat height shall be adjustable or suitable for safe transfer from the supplied beds.					
	4	Provide a removable covered sanitary pan.					
	5	Seat, backrest and arm supports shall be nonporous and easy to disinfect.					
	6	Provide non-slip feet or lockable castors for a mobile design.					
	7	Provide stable armrests for sitting and standing support.					
	8	All edges shall be rounded and free of sharp projections.					
	9	Supply one spare compatible sanitary pan.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
10		Demonstrate stability and safe pan removal.					
11		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
12		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
13		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
14		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
15		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
16		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
17		Inspect assembly, stability and surface condition at delivery before use.					
18		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
19		Patient support and mattress shall fit together without unsafe gaps and accommodate the patient group and load capacity stated for this item; provide a dimensioned drawing.					
20		Movable sections shall have protected mechanisms and safe clearances to prevent patient or operator entrapment during normal adjustment.					
21		All mobile equipment shall have hospital-grade casters with effective brakes on at least two wheels and shall roll safely over typical indoor door thresholds.					
22		Upholstery or mattress covers shall be fluid resistant, intact and readily disinfected.					
23		Patient transfer shall be possible using the supplied handles/rails and standard clinical staff technique without removing safety components.					
24		Demonstrate safe entry/exit or transfer and locking under the declared working conditions.					
25		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
26		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	27	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	28	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Three-function manual hospital bed							
EQ-100	100						
	1	Adult hospital bed with corrosion-resistant steel frame and smooth disinfectant-resistant finish.					
	2	Safe working load at least 180 kg, including patient and accessories.					
	3	Provide adjustable backrest and knee/leg section.					
	4	Provide secure, collapsible side rails with safe gaps and protected pinch points.					
	5	Removable head and foot boards for clinical access.					
	6	Supply a pressure-distributing mattress with removable waterproof cover and sealed or protected seams.					
	7	Bed and mattress dimensions shall be compatible; declare overall dimensions and usable sleeping area.					
	8	Four castors with at least two effective brakes.					
	9	Supply one IV pole and urine-bag hooks.					
	10	Demonstrate positioning, rail locking, brake function and safe working load documentation.					
	11	Three independent manual functions: backrest, knee/leg section and bed height.					
	12	Cranks shall fold away and have overload/end-stop protection.					
	13	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
14		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
15		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
16		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
17		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
18		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
19		Inspect assembly, stability and surface condition at delivery before use.					
20		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
21		Patient support and mattress shall fit together without unsafe gaps and accommodate the patient group and load capacity stated for this item; provide a dimensioned drawing.					
22		Movable sections shall have protected mechanisms and safe clearances to prevent patient or operator entrapment during normal adjustment.					
23		All mobile equipment shall have hospital-grade casters with effective brakes on at least two wheels and shall roll safely over typical indoor door thresholds.					
24		Upholstery or mattress covers shall be fluid resistant, intact and readily disinfected.					
25		Patient transfer shall be possible using the supplied handles/rails and standard clinical staff technique without removing safety components.					
26		Demonstrate safe entry/exit or transfer and locking under the declared working conditions.					
27		All functions specified for the bed type (backrest, knee, height and tilt as applicable) shall be independently operable through their full manufacturer-rated range, without binding under rated load.					
28		The specified assembly shall accommodate the intended patient/load and fit the approved installation or storage space; submit dimensioned manufacturer evidence before award.					
29		Provide protected controls or crank storage to limit accidental activation.					
30		The bed shall have a quick manual CPR release to flatten the backrest without mains power.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	31	Provide secure side-rail latches where rails are included.					
	32	The described component shall be user-serviceable or replaceable by the authorized service team, with the necessary service information and five-year availability in Yemen.					
	33	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	34	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	35	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	36	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

LOT 10 Sterilization and decontamination**Double- or triple-basin instrument-washing sink**

EQ-112	112						
	1	Freestanding sink fabricated from corrosion-resistant stainless steel, minimum grade 304.					
	2	Smooth welded joints, radiused internal corners and no exposed sharp edges.					
	3	Each basin shall have a separate drain, removable strainer and serviceable trap.					
	4	Provide splashback and sealed wall interface.					
	5	Provide adjustable feet and a rigid supporting frame.					
	6	Include mixing taps and all water/drain fittings within the agreed installation boundary.					
	7	Declare steel thickness; proposed basin/worktop minimum thickness 1.2 mm.					
	8	Fabrication dimensions shall be approved against room measurements before manufacture.					
	9	Demonstrate watertightness, drainage and resistance to rocking under working load.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
10		Provide three basins for sequential decontamination tasks and an adjacent draining area.					
11		Proposed overall length 1,800–2,100 mm and depth 600–750 mm; confirm workflow and room fit.					
12		The sink is for instrument processing and shall be separate from handwashing.					
13		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
14		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
15		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
16		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
17		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
18		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
19		Inspect assembly, stability and surface condition at delivery before use.					
20		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
21		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
22		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
23		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
24		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
25		Quote the offered basin count, stainless-steel grade, bowl dimensions, splashback, taps, drainage and clean/dirty workflow layout before fabrication.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
Stainless-steel decontamination sink, three-bowl							
EQ-115	115						
	1	Freestanding sink fabricated from corrosion-resistant stainless steel, minimum grade 304.					
	2	Smooth welded joints, radiused internal corners and no exposed sharp edges.					
	3	Each basin shall have a separate drain, removable strainer and serviceable trap.					
	4	Provide splashback and sealed wall interface.					
	5	Provide adjustable feet and a rigid supporting frame.					
	6	Include mixing taps and all water/drain fittings within the agreed installation boundary.					
	7	Declare steel thickness; proposed basin/worktop minimum thickness 1.2 mm.					
	8	Fabrication dimensions shall be approved against room measurements before manufacture.					
	9	Demonstrate watertightness, drainage and resistance to rocking under working load.					
	10	Provide three basins for sequential decontamination tasks and an adjacent draining area.					
	11	Proposed overall length 1,800–2,100 mm and depth 600–750 mm; confirm workflow and room fit.					
	12	The sink is for instrument processing and shall be separate from handwashing.					
	13	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	14	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	15	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
	16	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
17		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
18		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
19		Inspect assembly, stability and surface condition at delivery before use.					
20		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
21		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
22		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
23		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
24		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

Sterile storage shelving							
EQ-116	116						
	1	One shelving unit for protected storage of sterilized instrument packs.					
	2	Corrosion-resistant stainless steel or equivalent nonporous cleanable construction.					
	3	At least four adjustable shelves.					
	4	Proposed overall dimensions approximately 1,200 mm wide × 450 mm deep × 1,800 mm high, subject to room fit.					
	5	Lowest shelf shall clear the floor by at least 200 mm for cleaning.					
	6	Shelf edges shall not puncture sterile packaging.					
	7	Declared capacity at least 50 kg per shelf under evenly distributed load.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
	8	Provide stable feet and anti-tip fixing where needed.					
	9	Place within a clean controlled storage zone away from sinks and splash.					
	10	Confirm shelving-unit dimensions and sterile-pack storage volume before fabrication.					
	11	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	12	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	13	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
	14	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
	15	All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
	16	Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
	17	Inspect assembly, stability and surface condition at delivery before use.					
	18	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	19	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	20	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	21	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	22	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

LOT 11 Clinical furniture and patient transfer**Dual-function wheelchair**



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
1		Adult manual wheelchair with corrosion-resistant frame.					
2		Safe working load at least 120 kg.					
3		Seat width approximately 450–500 mm, subject to patient population and doorway clearance.					
4		Seat and backrest shall have washable durable upholstery.					
5		Provide effective wheel brakes accessible to the user or attendant.					
6		Provide removable or swing-away footrests and heel/calf supports.					
7		Provide armrests allowing safe transfer; removable or flip-up type preferred.					
8		Large rear wheels with hand rims and front swivel castors.					
9		Provide anti-tip features appropriate to the design.					
10		Provide folding frame where needed for transport and storage.					
11		Demonstrate brakes, folding mechanism and stability.					
12		Proposed dual function: wheelchair plus commode, with removable sanitary pan and a covered seat opening.					
13		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
14		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
15		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
16		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
17		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
	18	Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
	19	Inspect assembly, stability and surface condition at delivery before use.					
	20	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	21	Patient support and mattress shall fit together without unsafe gaps and accommodate the patient group and load capacity stated for this item; provide a dimensioned drawing.					
	22	Movable sections shall have protected mechanisms and safe clearances to prevent patient or operator entrapment during normal adjustment.					
	23	All mobile equipment shall have hospital-grade casters with effective brakes on at least two wheels and shall roll safely over typical indoor door thresholds.					
	24	Upholstery or mattress covers shall be fluid resistant, intact and readily disinfected.					
	25	Patient transfer shall be possible using the supplied handles/rails and standard clinical staff technique without removing safety components.					
	26	Demonstrate safe entry/exit or transfer and locking under the declared working conditions.					
	27	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	28	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	29	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	30	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Emergency Trolley with Basic Resuscitation Set							
EQ-120	120						
	1	Mobile emergency trolley with at least five drawers and a washable work surface.					
	2	Central drawer locking or tamper-evident breakaway sealing system.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
3		Clearly labelled drawer compartments for organized emergency supplies.					
4		Raised worktop edges and smooth disinfectant-resistant surfaces.					
5		Four castors with at least two brakes.					
6		Provide a defibrillator shelf compatible with the separately scheduled defibrillator or the existing unit.					
7		Provide a rigid CPR board.					
8		Provide IV pole, oxygen-cylinder holder and sharps-container bracket.					
9		Overall dimensions and loaded stability shall suit the receiving ward and doorways.					
10		Supply 20 numbered tamper seals if the offered locking system uses them.					
11		Trolley supply excludes medicines and separately scheduled medical devices; quote all trolley accessories in the unit price.					
12		Demonstrate drawer access, brakes and accessory retention under the declared safe load.					
13		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
14		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
15		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
16		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
17		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
18		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
19		Inspect assembly, stability and surface condition at delivery before use.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origon
20		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
21		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
22		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
23		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
24		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
25		Include a lockable emergency trolley, CPR board, IV pole, oxygen-cylinder holder, adult self-inflating resuscitation bag with masks and a labelled starter set of basic airways.					
26		List all medicines and consumables separately as an approved hospital checklist; do not assume drugs are included in the equipment price.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

Examination couch / emergency treatment bed

EQ-121	121						
	1	Clinical examination couch with rigid corrosion-resistant frame.					
	2	Patient surface shall have a seamless or sealed waterproof cover resistant to routine disinfectants.					
	3	Provide padded mattress or upholstery securely attached to the frame.					
	4	All edges and corners shall be smooth.					
	5	Provide stable non-slip feet or lockable castors if a mobile design is offered.					
	6	The item shall fit the receiving room and route of access; its load rating shall meet the item-specific minimum stated above, with dimensioned manufacturer drawing submitted for verification.					
	7	Supply paper-roll holder where compatible with the intended use.					
	8	Demonstrate stability and cleaning access.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
9		Adult patient surface approximately 1,800–1,900 mm long and at least 600 mm wide.					
10		Safe working load at least 180 kg.					
11		Provide adjustable backrest and a non-slip bedside step.					
12		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
13		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
14		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
15		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
16		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
17		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
18		Inspect assembly, stability and surface condition at delivery before use.					
19		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
20		Patient support and mattress shall fit together without unsafe gaps and accommodate the patient group and load capacity stated for this item; provide a dimensioned drawing.					
21		Movable sections shall have protected mechanisms and safe clearances to prevent patient or operator entrapment during normal adjustment.					
22		All mobile equipment shall have hospital-grade casters with effective brakes on at least two wheels and shall roll safely over typical indoor door thresholds.					
23		Upholstery or mattress covers shall be fluid resistant, intact and readily disinfected.					
24		Patient transfer shall be possible using the supplied handles/rails and standard clinical staff technique without removing safety components.					
25		Demonstrate safe entry/exit or transfer and locking under the declared working conditions.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
	26	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	27	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	28	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	29	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Instrument Trolley, Stainless Steel							
EQ-123	123						
	1	Hospital mobile trolley with corrosion-resistant frame and smooth disinfectant-resistant surfaces.					
	2	Provide four castors with at least two brakes.					
	3	Provide rounded edges, push handle and stable construction.					
	4	Supply tray/shelf retaining edges appropriate to the contents.					
	5	The item shall fit the receiving room and route of access; its load rating shall meet the item-specific minimum stated above, with dimensioned manufacturer drawing submitted for verification.					
	6	Proposed footprint approximately 600–800 mm wide × 400–550 mm deep; confirm room and doorway fit.					
	7	Components shall tolerate repeated cleaning without trapping liquid or debris.					
	8	Demonstrate drawer/door function where fitted and loaded trolley stability.					
	9	Surgical instrument configuration with two removable stainless-steel trays/shelves.					
	10	Provide sufficient clearance between shelves for covered instrument trays.					
	11	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
12		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
13		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
14		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
15		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
16		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
17		Inspect assembly, stability and surface condition at delivery before use.					
18		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
19		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
20		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
21		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
22		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					

TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply

Medicine and Instrument Cabinet

EQ-124	124						
1		Closed storage cabinet with smooth nonporous cleanable surfaces.					
2		Corrosion-resistant construction; use stainless steel minimum grade 304 for sterile-instrument storage.					
3		Provide at least three adjustable shelves.					
4		Doors shall close securely and include a lock with two keys where access control is needed.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
5		Provide stable feet or plinth that allows cleaning and prevents tipping.					
6		Shelf surfaces shall not damage stored packs.					
7		Proposed overall dimensions approximately 900 mm wide × 450 mm deep × 1,800 mm high.					
8		Provide a declared shelf-load capacity of at least 30 kg per shelf.					
9		Confirm cabinet dimensions and door swing against the room layout before fabrication.					
10		Medicines or instruments requiring controlled temperature shall not be stored on the assumption that this cabinet provides it.					
11		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
12		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
13		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
14		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
15		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
16		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
17		Inspect assembly, stability and surface condition at delivery before use.					
18		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
19		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
20		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
21		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	22	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Mobile bedside privacy screen							
EQ-125	125						
	1	Mobile privacy screen with at least three hinged panels.					
	2	Overall extended width at least 1,800 mm and height at least 1,650 mm.					
	3	Frame shall be corrosion resistant with rounded edges.					
	4	Panels shall be opaque and washable or removable for laundering.					
	5	Provide stable feet/castors with brakes where fitted.					
	6	Screen shall fold for storage without pinch hazards.					
	7	Materials shall meet the facility's applicable fire-performance requirements.					
	8	Supply all panels and frame fixings as one complete unit.					
	9	Demonstrate stability, folding and cleaning access.					
	10	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	11	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	12	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
	13	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
	14	All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	15	Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
	16	Inspect assembly, stability and surface condition at delivery before use.					
	17	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	18	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	19	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	20	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	21	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

LOT 13 Laundry infection prevention and safety**Clinical Handwashing Station / Sink**

EQ-132	132						
	1	Wall-mounted clinical handwashing sink with smooth nonporous basin and splash-resistant surrounding surfaces.					
	2	Hands-free tap using elbow, knee, foot or sensor control suitable for local maintenance.					
	3	Water outlet shall not discharge directly into the drain opening.					
	4	Provide hot/cold mixing only where the approved water service supports it.					
	5	Include drainage trap, isolating valves, connection hoses and sealed wall fixings.					
	6	Supply wall-mounted liquid-soap dispenser and disposable-towel dispenser.					
	7	Provide an adjacent pedal-operated waste bin.					
	8	Sink shall be separate from instrument decontamination sinks.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
	9	Agree basin dimensions, tap reach and mounting height with the hospital before fabrication.					
	10	Demonstrate leak-free drainage and reliable hands-free operation after installation.					
	11	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	12	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	13	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
	14	Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
	15	All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
	16	Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
	17	Inspect assembly, stability and surface condition at delivery before use.					
	18	Demonstrate all adjustments, locks and brakes fitted to the offered item.					
	19	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	20	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	21	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	22	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Color-coded reusable laundry bags and stands							
EQ-134	134						
	1	Each unit shall consist of one reusable colour-coded laundry bag and one compatible mobile/freestanding holder.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
2		Proposed usable bag capacity 60–100 L.					
3		Bag fabric shall withstand the hospital's validated laundry temperature and chemicals.					
4		Provide secure bag closure and lifting handles appropriate to the rated load.					
5		Holder shall be corrosion resistant with stable base and hands-free lid where specified.					
6		Provide castors with brakes on mobile holders.					
7		Colours and labels shall match the hospital's dirty-linen segregation policy.					
8		Include one spare bag of the same type per holder.					
9		The supplied device shall implement any limitations on wet or infectious linen; provide leak-resistant liners where required by the local protocol as an enabled safety function, with audible/visual indication where relevant and a documented simulated-fault acceptance test.					
10		Confirm that the scheduled quantity of ten means ten bag-and-stand sets.					
11		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
12		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
13		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
14		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
15		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
16		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
17		Inspect assembly, stability and surface condition at delivery before use.					
18		Demonstrate all adjustments, locks and brakes fitted to the offered item.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
	19	Provide exact catalogue reference and unit-of-issue for every supplied component.					
	20	Patient- and chemical-contact materials shall be compatible with the specified clinical use and cleaning/disinfection method and shall be free from prohibited latex where direct contact is expected.					
	21	Provide manufacturer labels showing lot/batch traceability where applicable.					
	22	Expiring supplies shall be delivered with at least 75% of labelled shelf life remaining and storage conditions compatible with the hospital supply chain.					
	23	All reusable components shall have validated cleaning/reprocessing instructions; all single-use components shall be labelled as such and supplied in the quantities specified for this item.					
	24	Inspect quantities, packaging integrity and component identity at receipt.					
	25	Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
	26	The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
	27	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	28	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Double-basin soaking and stain-treatment sink							
EQ-136	136						
	1	Freestanding sink fabricated from corrosion-resistant stainless steel, minimum grade 304.					
	2	Smooth welded joints, radiused internal corners and no exposed sharp edges.					
	3	Each basin shall have a separate drain, removable strainer and serviceable trap.					
	4	Provide splashback and sealed wall interface.					
	5	Provide adjustable feet and a rigid supporting frame.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
6		Include mixing taps and all water/drain fittings within the agreed installation boundary.					
7		Declare steel thickness; proposed basin/worktop minimum thickness 1.2 mm.					
8		Fabrication dimensions shall be approved against room measurements before manufacture.					
9		Demonstrate watertightness, drainage and resistance to rocking under working load.					
10		Provide two basins for soaking and stain treatment of linen.					
11		Proposed overall length 1,400–1,800 mm and depth 650–750 mm; confirm room fit.					
12		Provide chemical-resistant basin surfaces compatible with approved laundry agents.					
13		The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
14		Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
15		Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					
16		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
17		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
18		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
19		Inspect assembly, stability and surface condition at delivery before use.					
20		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
21		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
22		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origion
	23	Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
	24	Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Stainless-steel soiled-linen sorting table							
EQ-138	138						
	1	Stainless-steel table dedicated to soiled-linen sorting.					
	2	Minimum grade 304 stainless-steel worktop with smooth welded joints.					
	3	Proposed worktop size approximately 1,500 mm × 750 mm and working height 850–900 mm.					
	4	Worktop thickness at least 1.2 mm with adequate stiffening underneath.					
	5	Provide raised rear/side edges where compatible with workflow.					
	6	Provide stable adjustable feet or lockable castors as approved for the room.					
	7	Safe evenly distributed worktop load at least 100 kg.					
	8	All surfaces shall be accessible for cleaning and free from sharp edges.					
	9	Keep the table in the dirty-linen zone with no crossover to clean-linen storage.					
	10	Confirm final dimensions and drainage/cleaning arrangements before fabrication.					
	11	The item shall fit the approved room and doorway dimensions; submit a manufacturer dimensioned drawing and weight so the purchaser can verify placement before award.					
	12	Patient-facing metal surfaces shall be smooth, corrosion-resistant and cleanable; stainless steel components shall be at least grade 304 where this grade is specified for the item.					
	13	Load-bearing surfaces shall safely support the item-specific minimum load stated in this schedule without permanent deformation during acceptance testing.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
14		Edges, welds and fasteners shall be smooth, secure and accessible for cleaning.					
15		All exposed surfaces shall tolerate routine hospital detergent and at least one commonly used healthcare disinfectant at the manufacturer-approved concentration without corrosion or damage.					
16		Moving or wearing parts shall be individually replaceable through an authorized Yemen service route for at least five years after acceptance.					
17		Inspect assembly, stability and surface condition at delivery before use.					
18		Demonstrate all adjustments, locks and brakes fitted to the offered item.					
19		Provide exact manufacturer, model/version and country of manufacture, evidenced by manufacturer catalogue and product label; no unidentifiable or mixed-model substitute is acceptable.					
20		The quoted price shall include all item-specific accessories, installation and commissioning stated here; any site infrastructure connection shall be listed and coordinated before contract award.					
21		Supply a five-year spare/accessory/consumable price list and evidence of an authorized service/replenishment route to Yemen; identify lead times, without reducing the mandatory base package.					
22		Provide a signed delivery/acceptance checklist matching this item's specifications and included components.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							

LOT 15 Basic doctors office and waiting-area furniture

EQ-143	143	Doctors office desk, approximately 1200 x 600 x 750 mm					
1		Provide a simple office desk for doctors documentation/consultation work; dimensions approximately 1200 W x 600 D x 750 H mm, tolerance ±50 mm.					
2		Worktop shall be at least 25 mm moisture-resistant laminated board or equivalent durable easy-clean construction.					
3		Provide smooth sealed edges with impact-resistant edging at least 2 mm thick or equivalent; sharp corners and exposed staples are unacceptable.					
4		Support frame shall be powder-coated steel or equivalent rigid construction, with stable joints and no perceptible rocking under normal use.					
5		Provide at least two drawers including one lockable drawer; supply two keys for each lock.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	6	Provide clear leg space at least 600 mm wide, 450 mm deep and 650 mm high under the desk.					
	7	Desk shall support at least 50 kg evenly distributed on the worktop without permanent deformation.					
	8	Provide adjustable or non-marking protective feet appropriate to the hospital floor.					
	9	Provide a cable access grommet or protected rear cable route suitable for a computer and printer.					
	10	All accessible surfaces shall permit routine wiping with manufacturer-approved hospital cleaning agents; untreated absorbent edges are unacceptable.					
	11	Use a neutral finish such as light grey or light wood effect; colour shall be coordinated with the receiving room.					
	12	Deliver fully assembled, positioned and levelled at the designated doctors office; include all hardware and drawer runners.					
	13	Demonstrate stable support, smooth drawer movement and functioning locks; replace cracked, swollen or scratched panels before acceptance.					
	14	Provide at least 12 months warranty against manufacturing, joint and drawer-runner defects, with local repair support.					
	15	Proposed quantity: two desks per hospital; final room allocation and quantity require purchaser confirmation before order.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Doctors adjustable office chair							
EQ-144	144						
	1	Provide an adjustable swivel office chair for doctors administrative use, one per proposed desk.					
	2	Rated user load shall be at least 120 kg, supported by manufacturer data or relevant load-test evidence.					
	3	Seat height shall be gas-lift adjustable over at least approximately 430–550 mm.					
	4	Provide a stable five-leg base with five free-running castors suitable for the hospital floor.					
	5	Provide a padded seat approximately 450 mm wide x 430 mm deep or larger.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	6	Provide an ergonomically contoured backrest with lumbar support and adjustable tilt or back inclination with locking.					
	7	Provide two secure armrests with smooth rounded contact surfaces.					
	8	Seat and back covering shall be smooth wipe-clean non-absorbent vinyl/PU or equivalent; absorbent fabric/mesh is not accepted for this specified easy-clean configuration.					
	9	Covering, seams and upholstery shall withstand the manufacturer-approved routine cleaning method without cracking or liquid ingress.					
	10	Frame and base shall be corrosion-resistant metal or reinforced material suitable for the rated user load.					
	11	Gas lift and tilt controls shall be accessible, with no exposed sharp edges, pinch hazards or uncontrolled movement in normal adjustment.					
	12	Provide complete assembly, height/tilt adjustment demonstration and castor inspection at delivery.					
	13	Demonstrate stable sitting, swivel and height retention at representative load; reject wobbling joints or sinking lifts.					
	14	Provide at least 12 months warranty and local replacement availability for castors and gas lift.					
	15	Proposed quantity: two chairs per hospital; final quantity requires purchaser confirmation.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Doctors office visitor chair, fixed frame							
EQ-145	145						
	1	Provide a simple fixed-frame visitor chair for doctors offices/consultation rooms.					
	2	Rated user load shall be at least 120 kg per chair, supported by manufacturer data or test evidence.					
	3	Seat height approximately 430–480 mm; seat width at least 450 mm and usable depth at least 400 mm.					
	4	Provide a welded or securely bolted corrosion-resistant steel frame with powder-coated or equivalent durable finish.					
	5	Use a four-leg or stable cantilever design, without castors or height adjustment.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	6	Provide a comfortable supportive seat and back with rounded edges and no accessible sharp parts.					
	7	Contact surfaces shall be non-absorbent wipe-clean molded material or sealed vinyl upholstery; absorbent fabric is not accepted.					
	8	Where padded, upholstery shall be firmly secured and resist routine cleaning without cracking or exposing foam.					
	9	Provide non-slip, non-marking floor protectors and stable seating without rocking.					
	10	Armrests, if present, shall be securely attached and easily cleaned without inaccessible dirt traps.					
	11	Chairs shall permit routine cleaning of the underside, back and frame with manufacturer-approved agents.					
	12	Deliver fully assembled; stacking is desirable where it does not damage the seat surfaces.					
	13	Inspect frame joints, surface finish, upholstery and stability at acceptance; damaged or unstable units shall be replaced.					
	14	Provide at least 12 months manufacturing-defect warranty and local replacement feet/parts.					
	15	Proposed quantity: four chairs per hospital, allowing two visitors per doctors desk; final quantity requires purchaser confirmation.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
EQ-146	146	Waiting-area bench, three seats per unit					
	1	Supply a simple three-seat connected waiting bench; one BOQ unit means one complete three-seat bench, not three separate chairs.					
	2	Overall width approximately 1600–1900 mm; depth approximately 550–750 mm; seat height approximately 430–480 mm.					
	3	Rated load shall be at least 120 kg per seat, 360 kg total per three-seat bench, with supporting manufacturer data or test evidence.					
	4	Frame and beam shall be corrosion-resistant powder-coated or stainless steel with secure seat supports.					
	5	Seats/backrests shall be molded easy-clean non-absorbent material or perforated metal with a durable smooth finish.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	6	Provide rounded seat edges and safe perforations; sharp edges, exposed burrs and pinch points are unacceptable.					
	7	Provide stable floor supports with non-slip/non-marking adjustable feet or equivalent levelling arrangement.					
	8	Provide armrests at both ends; fixings shall remain secure under normal repeated patient use.					
	9	Seat/back assemblies shall be securely attached and individually replaceable where practical.					
	10	All surfaces shall be accessible for routine cleaning; untreated wood and absorbent fabric are not accepted.					
	11	Select a layout that retains clear circulation, emergency egress and wheelchair access at the agreed waiting area.					
	12	Supply all assembly and floor-fixing hardware where required by the approved layout; do not drill floors without site agreement.					
	13	Demonstrate stability, joint security and load-rating evidence at acceptance; replace distorted or damaged parts.					
	14	Provide at least 12 months manufacturing-defect warranty and local repair support for joints, feet and seating panels.					
	15	Proposed allocation: two benches per hospital, providing six seats per hospital; final quantity and layout require purchaser confirmation.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							
Lockable doctors office filing cabinet							
EQ-147	147						
	1	Supply one upright lockable cabinet for doctors files and office supplies; it is not a medicine or sterile-instrument storage cabinet.					
	2	Approximate dimensions 900 W x 450 D x 1800 H mm; tolerance ± 100 mm, subject to the verified office doorway and space.					
	3	Body and doors shall be corrosion-protected powder-coated steel, nominal sheet thickness at least 0.7 mm or equivalent rigidity.					
	4	Provide at least four adjustable internal shelves plus the cabinet base.					
	5	Each shelf shall support at least 30 kg evenly distributed without permanent deformation.					



Item ID	No	Furniture / Detailed technical requirement	Catalogue page / evidence	Supplier Technical Specification	Model	Brand	Country of Origin
	6	Provide two hinged doors or sliding doors that operate smoothly without obstructing the approved circulation space.					
	7	Provide a secure lock with at least two keys, suitable for confidential patient-related files.					
	8	Provide smooth rounded/folded edges; exposed sharp sheet-metal edges or unstable door hinges are unacceptable.					
	9	Supply an anti-tip restraint and suitable fixing hardware; install it to a verified suitable wall/substrate before loading.					
	10	Provide stable levelling feet/base and non-marking floor protection.					
	11	Interior and exterior surfaces shall permit routine wiping; finish shall be continuous and free of corrosion, flaking paint and deep scratches.					
	12	Supply fully assembled, positioned and secured, with all shelf clips and door hardware included.					
	13	Verify stability, anti-tip installation, shelf engagement, lock and door operation at acceptance.					
	14	Provide at least 12 months manufacturing-defect warranty and local availability of locks, keys and shelf supports.					
	15	Proposed allocation: one cabinet per hospital; confirm final quantity and room placement before order.					
TOTAL MAXIMUM SCORE — item is eligible only if all Mandatory clauses comply							